

vent.

Nashoba Analytical, LLC

Tel: 978-391-4428 Fax: 978-391-4643

31 A Willow Rd, Ayer MA 01432

Web Site <http://www.nashobanalytical.com>

Attention:
Darren MacCaughey
Town of Westford
55 Main Street
Westford, MA 01886

RECEIVED

Report Date: 5/21/2013

MAY 21 2013

WESTFORD BOARD OF HEALTH

Date Sampled: 5/20/2013

Date Received: 5/20/2013

Date of Analysis: 5/20/2013

Summary Report

Sample Location	E.coli per 100 ml	Lab#
1 Nashoba Valley Function Pond	ND	136671
2 Nashoba Valley Campers Beach	ND	136671
3 Summer Village Main Beach	ND	136672
4 American Legion at LISP	ND	136673
5 Lakeside Meadows (on Nab Lake)	8	136676
6 North Beach (on Nab Lake) N.I.A.	34	136677
7 Mary Lou Beach (on Nab Lake) N.I.A.	6	136677
8 Noble Cove (on Nab Lake) N.I.A.	2	136677
9 Sandy Beach (on Nab Lake) N.I.A.	ND	136677
10 East Boston Camps (Daycare Beach)	ND	136678
11 East Boston Camps (Boys Beach)	ND	136678
12 East Boston Camps (Girls Beach)	2	136678

NI=Not Tested

ND=None Detected

DPH Limit = 235 cfu/100ml
Method of Analysis: EPA 1603

Massachusetts State Certified
Laboratory #MA1118


David L. Knowlton
Laboratory Director