

Attention:
Jeff Stephens/Rae Dick
Town of Westford
55 Main Street
Westford, MA 01886

Report Date: 8/1/2018

Date Sampled: 7/31/2018

Date Received: 7/31/2018

Date of Analysis: 7/31/2018

Summary Report

Sample Location	E.coli per 100 ml	Lab#
1 Nashoba Valley Campers Beach	ND	192552
2 Nashoba Valley Tiki Pond	4	192552
3 Forge Village Beach	20	192554
4 Summer Village Main Beach	14	192555
5 American Legion at LSFP	10	192556
6 Lakeside Meadows (on Nab Lake)	ND	192558
7 North Beach (on Nab Lake) N.I.A.	ND	192559
8 Mary Lou Beach (on Nab Lake) N.I.A.	ND	192559
9 Edwards Beach	ND	192554
10 Sandy Beach (on Nab Lake) N.I.A.	44	192559
11 Noble Cove (on Nab Lake) N.I.A.	30	192559
12 East Boston Camps (Boys Beach)	4	192560
13 East Boston Camps (Girls Beach)	6	192560
14 East Boston Camps (Day Camp Beach)	ND	192560

NT=Not Tested
ND=None Detected


DPH Limit = 235 cfu/100ml
Method of Analysis: EPA 1603

Massachusetts State Certified
Laboratory # = M-MA1118

RECEIVED

AUG 3 2018

WESTFORD
BOARD OF HEALTH


David L. Knowlton
Laboratory Director